



Certificate of Conformity

Number NMI-CoC-25.05 revision 0
Project number 3851894
Page 1 of 3

Issued by : NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands

Applicant : CET Electric Technology Inc.
Floor 33-35, Building #6
Shenzhen International Innovation Valley
Dashi First Road, Nanshan District, Shenzhen
Guangdong, 518055
P.R. China

Submitted : **Power Quality Analyzer**

Manufacturer : CET Electric Technology Inc.
Type : iMeter 5
Destined for the : Electrical Power quality meter
measurement of

In accordance with : See page 2

Characteristics : See page 3

The undersigned declares that the described product is tested according to the standards as referred to on page 2 and meets their requirements, based on a non-recurrent examination. The appertaining test data is presented in the type evaluation report NMI-3851894-01, granted by NMI.

NMI Certin B.V.
8 April 2025

Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.



IEC 61000-4-30 Power Quality functions tested

The following IEC 61000-4-30 measurement methods have been tested

Table 1 IEC 61000-4-30 Power Quality functions tested

IEC 62586-2 Clause	Parameter	IEC 61000-4-30 class	Comments
6.1 / 7.1	Power frequency	A + S	50 and 60 Hz
6.2 / 7.2	Magnitude of supply voltage	A + S	
6.3 / 7.3	Flicker	A + S	Class F1 230 V 50 Hz, 60 Hz
6.4 / 7.4	Supply voltage interruptions, dips and swells	A + S	50 and 60 Hz
6.5 / 7.5	Supply voltage unbalance	A + S	
6.6 / 7.6	Voltage harmonics	A + S	
6.7 / 7.7	Voltage interharmonics	A + S	
6.8 / 7.8	Mains signalling voltages on the voltage supply	A + S	Method 1
6.9 / 7.9	Measurement of underdeviation and overdeviation parameters	A + S	
6.10 / 7.10	Flagging	A + S	
6.11 / 7.11	Clock uncertainty testing	A + S	
6.12 / 7.12	Variation of external influence quantities	A + S	Temperature: -25°C to + 70°C
6.13 / 7.13	Rapid Voltage Changes (RVC)	A + S	
6.14 / 7.14	Magnitude of current	A + S	
6.15 / 7.15	Harmonic current	A + S	
6.16 / 7.16	Interharmonic currents	A + S	
6.17 / 7.17	Current unbalance	A + S	
8	Calculation of measurement uncertainty and operating uncertainty	A + S	
<p>A : compliance with class A S : compliance with class S --- : Not implemented</p> <p>The tests are performed in accordance with IEC 62586-2.</p>			

Characteristics of the measuring instrument

Model	iMeter 5
Accuracy class	Class A and Class S
U_{ref}	230 VLN
I_{test}	5 A
I_{max}	20 A
f_{ref}	50 Hz and 60 Hz
Temperature	Rated range of operation: -25 °C to +70°C
Power supply range	95-305VAC/VDC \pm 10%, 47-440 Hz
Firmware measure version	V1.50.00
Software version	V1.50.00
Hardware version	V1

Certificate history:

Revision	Date	Description of the modification
0	8 April 2025	First issue